



Surgical Aspirator Tip (SS) Cleaning & Sterilization Instructions

1. Place the Surgical Aspirator Tip into the ultrasonic cleaner or instrument washer
2. Run complete cycle based upon machine manufacturer's instructions.
3. Rinse device thoroughly to completely remove cleaning solution from the device.
4. Bag instrument and sterilized based upon the sterilizer manufacturer's instructions.