



Uni-Tip Cleaning & Sterilization Instructions

1. Place Uni-Tip air/water syringe tips in ultrasonic cleaner or instrument washer
2. Run complete cycle based upon machine manufacturer's instructions.
3. Rinse device thoroughly to completely remove cleaning solution from the device.
4. Bag instrument and sterilized based upon the sterilizer manufacturer's instructions.